



Model 601 CO Interference Data

Interference Response

Date of Test 6/28/2006
Analyzer Type CO
Model No. 602
Serial No. T06034-M
Calibration Span 3000ppm

Test Gas Type		Concentration	Analyzer Response	
		(ppm)	Wet	Dry
H2O		2.5%	-5	0
CO2		5%	-6	-1
CO2		15%	-7	-2
CO		N/A	N/A	N/A
CH4		50	-5	0
SO2		20	0	0
NH3		15	0	0
NO		13	-5	0
N2O		11	-3	2
NO2		9	-4	2