



Recommended Spare Parts – 2014

Model 600 HCLD Gas Analyzer

Part Number	Description
110386-1	EPC Valve (Air)
110271-2	EPC Valve (Sample)
110773-3	EPC Valve Rebuild Kit (Air)
110773-2	EPC Valve Rebuild Kit (Sample)
410289-2	Fuse, 8 Amp 250V (Slo Blo) (115 VAC)
410291-4	Fuse, 4 Amp 250V (Slo Blo) (230 VAC)
310127-1	Filter Element (Sample)
310074-1	Filter, Long Pass (Red Lens)
110378-1R	NO _x Converter
110083-1	NO _x Converter Carbon Refill Pack
410307-2	Ozone Lamp
410309-1	Ozone Lamp Power Supply
310821-1	Reaction Chamber
310081-1	Reaction Chamber O-Ring
410797-1	RTD, Cell
410797-2	RTD, Converter
410797-3	RTD, Pump
410798-1	RTD, Oven
110417-5	Heater, Cell 115V
110417-6	Heater, Cell 230V
110419-3	Heater, Converter 115V
110419-4	Heater, Converter 230V
110416-3	Heater, Pump 115V
110416-4	Heater, Pump 230V
110418-3	Heater, Oven 115V
110418-4	Heater, Oven 230V
310235-1	Regulator, Air
310005-7	Pump, Sample 115V (N012)
310005-8	Pump, Sample 230V (N012)
310309-8	Rebuild Kit, Pump N012
310309-7	Rebuild Kit, Pump N010
110568-2R	Nozzle, Sample (Standard)
110568-1R	Nozzle, Sample (Low Pressure)
110568-4R	Nozzle, Sample (5000 ppm)
110569-3R	Nozzle, O ₃ (Standard)
110569-1R	Nozzle, O ₃ (Ozone Pump)
410196-1	Sensor Pressure (0-5 psi) Sample
410196-2	Sensor Pressure (0-30 psi) Air
410196-3	Sensor Pressure (0-1 psi) Sample (Low Pressure)
110487-4	Solenoid Valve Assembly (Span) (Optional)
110487-3	Solenoid Valve Assembly (Zero) (Optional)

List subject to change without notice.

California Analytical Instruments

1312 West Grove Ave., Orange, CA 92865 • Phone: 714-974-5560 • Fax: 714-921-2531

www.gasanalyzers.com